

Certificate of Compliance

Certificate: 2257425

Master Contract: 248956

Project: 2257425

Date Issued: April 7, 2010

Issued to: Phoenix Analysis & Design Technologies, Inc.
7755 South Research Dr, 110
Tempe, AZ 85284
USA

Attention: Mr. Rey Chu

The products listed below are eligible to bear the CSA Mark shown



Issued by: Justin Liu



Authorized by: Anthony Lee P. Eng
Certification Specialist - Asia

PRODUCTS

CLASS -8721 05 - LABORATORY EQUIPMENT-Electrical

Support Cleaning Apparatus, Model SCA-1200, rated 100-120 Vac, 60 Hz, 12 A.

Note: The above model is Class I, Installation Category II equipment for indoor use in Pollution Degree 2 environment, 10 to 30 deg C ambient, and up to 2000 m.

APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

CAN/CSA-C22.2 No. 61010-2-010 - Part 2 Particular Requirements, Section 2.10 Specification for laboratory equipment for the heating of materials.



CSA INTERNATIONAL

Supplement to Certificate of Compliance

Certificate: 2257425

Master Contract: 248956

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
2257425	April 7, 2010	Original Certification.